

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX083120\
 Data File : VX018152.D
 Acq On : 31 Aug 2020 17:02
 Operator : JC/SP
 Sample : L3846-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB181-GW-738-740

Quant Time: Sep 01 06:34:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	489198	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	791473	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	733720	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	361159	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	340643	50.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.32%	
35) Dibromofluoromethane	5.47	113	271449	50.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.72%	
50) Toluene-d8	8.70	98	949027	47.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.54%	
62) 4-Bromofluorobenzene	11.12	95	352745	44.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.32%	
Target Compounds						
16) Acetone	2.42	43	9097	3.438	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX083120\
Data File : VX018152.D
Acq On : 31 Aug 2020 17:02
Operator : JC/SP
Sample : L3846-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB181-GW-738-740

Quant Time: Sep 01 06:34:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

