

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX083120\
 Data File : VX018153.D
 Acq On : 31 Aug 2020 17:25
 Operator : JC/SP
 Sample : L3835-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE BLANK-SOILREF-6

Quant Time: Sep 01 06:36:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	486193	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	794060	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	750750	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	353406	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	339979	50.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.74%	
35) Dibromofluoromethane	5.47	113	276338	51.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.20%	
50) Toluene-d8	8.70	98	960005	47.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.32%	
62) 4-Bromofluorobenzene	11.12	95	355807	44.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.80%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX083120\
Data File : VX018153.D
Acq On : 31 Aug 2020 17:25
Operator : JC/SP
Sample : L3835-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGEBLANK-SOILREF-6

Quant Time: Sep 01 06:36:28 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

