

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX083120\
 Data File : VX018156.D
 Acq On : 31 Aug 2020 18:33
 Operator : JC/SP
 Sample : L3835-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGEBLANK-SAMPLEMANAG

Quant Time: Sep 01 06:44:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	454261	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	741709	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	701810	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	347566	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	332207	52.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.36%	
35) Dibromofluoromethane	5.46	113	258050	51.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.16%	
50) Toluene-d8	8.70	98	904431	48.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.14%	
62) 4-Bromofluorobenzene	11.12	95	340565	46.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.02%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX083120\
Data File : VX018156.D
Acq On : 31 Aug 2020 18:33
Operator : JC/SP
Sample : L3835-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGEBLANK-SAMPLEMANAG

Quant Time: Sep 01 06:44:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

