

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX083120\
 Data File : VX018157.D
 Acq On : 31 Aug 2020 18:56
 Operator : JC/SP
 Sample : L3831-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GR-OF-20200827

Quant Time: Sep 01 06:47:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	446903	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	737726	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	683913	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	326701	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	323797	52.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.38%	
35) Dibromofluoromethane	5.47	113	253416	50.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.88%	
50) Toluene-d8	8.70	98	894885	47.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.64%	
62) 4-Bromofluorobenzene	11.12	95	324429	44.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.12%	
Target Compounds						
16) Acetone	2.42	43	10911	4.513	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX083120\
Data File : VX018157.D
Acq On : 31 Aug 2020 18:56
Operator : JC/SP
Sample : L3831-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GR-OF-20200827

Quant Time: Sep 01 06:47:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

