

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX083120\
 Data File : VX018166.D
 Acq On : 31 Aug 2020 23:30
 Operator : JC/SP
 Sample : PB131285TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB131285TB

Quant Time: Sep 01 07:15:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	464007	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	755468	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	711984	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	331201	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	334977	52.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.00%	
35) Dibromofluoromethane	5.47	113	254495	49.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.92%	
50) Toluene-d8	8.70	98	924051	48.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.44%	
62) 4-Bromofluorobenzene	11.12	95	336619	44.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.30%	
Target Compounds						
16) Acetone	2.42	43	13740	5.474	ug/l	95
20) Methylene Chloride	2.83	84	15763	1.949	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX083120\
Data File : VX018166.D
Acq On : 31 Aug 2020 23:30
Operator : JC/SP
Sample : PB131285TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB131285TB

Quant Time: Sep 01 07:15:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

