

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX083120\
 Data File : VX018168.D
 Acq On : 01 Sep 2020 00:15
 Operator : JC/SP
 Sample : PB131295TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB131295TB

Quant Time: Sep 01 07:19:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	470879	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	764918	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	736316	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	340475	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	336758	51.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.04%	
35) Dibromofluoromethane	5.47	113	263389	50.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.12%	
50) Toluene-d8	8.70	98	950724	49.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.00%	
62) 4-Bromofluorobenzene	11.12	95	352795	46.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.42%	
Target Compounds						
16) Acetone	2.42	43	47080	18.483	ug/l	89
18) Methyl Acetate	2.75	43	7526	1.224	ug/l #	87
20) Methylene Chloride	2.84	84	12997	1.359	ug/l	85
43) Isopropyl Acetate	6.42	43	126567	9.563	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX083120\
Data File : VX018168.D
Acq On : 01 Sep 2020 00:15
Operator : JC/SP
Sample : PB131295TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB131295TB

Quant Time: Sep 01 07:19:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

