

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX083120\
 Data File : VX018169.D
 Acq On : 01 Sep 2020 00:38
 Operator : JC/SP
 Sample : L3508-29
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TR-08-082820

Quant Time: Sep 01 07:23:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	464913	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	754313	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	714762	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	346518	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	329339	51.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.06%	
35) Dibromofluoromethane	5.47	113	257176	50.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.12%	
50) Toluene-d8	8.70	98	913284	47.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.46%	
62) 4-Bromofluorobenzene	11.12	95	342717	45.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.06%	
Target Compounds						
16) Acetone	2.42	43	12043	4.789	ug/l #	78
20) Methylene Chloride	2.83	84	12807	1.354	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX083120\
Data File : VX018169.D
Acq On : 01 Sep 2020 00:38
Operator : JC/SP
Sample : L3508-29
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TR-08-082820

Quant Time: Sep 01 07:23:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

