

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX083120\
 Data File : VX018172.D
 Acq On : 01 Sep 2020 01:47
 Operator : JC/SP
 Sample : L3828-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BS-2

Quant Time: Sep 01 07:30:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	446631	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	727159	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	715225	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	326758	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	325704	52.53	ug/l	0.00
Spiked Amount			Recovery	=	105.06%	
35) Dibromofluoromethane	5.46	113	252950	51.08	ug/l	0.00
Spiked Amount			Recovery	=	102.16%	
50) Toluene-d8	8.70	98	928855	50.36	ug/l	0.00
Spiked Amount			Recovery	=	100.72%	
62) 4-Bromofluorobenzene	11.12	95	336378	46.35	ug/l	0.00
Spiked Amount			Recovery	=	92.70%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	3.02	59	8715	7.229	ug/l	99
13) Acrolein	2.27	56	275	0.550	ug/l #	77
16) Acetone	2.42	43	57021	23.601	ug/l	94
18) Methyl Acetate	2.75	43	7465	1.280	ug/l #	83
20) Methylene Chloride	2.84	84	28645	4.741	ug/l	95
25) 2-Butanone	4.64	43	36977	9.313	ug/l	97
43) Isopropyl Acetate	6.42	43	129465	10.290	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX083120\
Data File : VX018172.D
Acq On : 01 Sep 2020 01:47
Operator : JC/SP
Sample : L3828-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BS-2

Quant Time: Sep 01 07:30:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

