

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX083120\
 Data File : VX018174.D
 Acq On : 01 Sep 2020 02:32
 Operator : JC/SP
 Sample : L3829-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 F06045

Quant Time: Sep 01 07:40:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	460818	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	734385	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	698813	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	339049	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	325276	50.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.70%	
35) Dibromofluoromethane	5.46	113	254336	50.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.70%	
50) Toluene-d8	8.70	98	896201	48.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.22%	
62) 4-Bromofluorobenzene	11.12	95	331533	45.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.46%	
Target Compounds						
16) Acetone	2.42	43	26800	10.751	ug/l	99
40) Benzene	6.12	78	11045	0.555	ug/l #	87
52) Toluene	8.76	92	16115	1.277	ug/l	95
68) m/p-Xylenes	10.34	106	4407	0.497	ug/l	92
95) Naphthalene	13.82	128	11765	0.525	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX083120\
Data File : VX018174.D
Acq On : 01 Sep 2020 02:32
Operator : JC/SP
Sample : L3829-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
F06045

Quant Time: Sep 01 07:40:13 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

