

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX083120\
 Data File : VX018175.D
 Acq On : 01 Sep 2020 02:55
 Operator : JC/SP
 Sample : L3854-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1332-FM648

Quant Time: Sep 01 07:54:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	472943	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	743868	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	714664	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	343305	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	331877	50.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.10%	
35) Dibromofluoromethane	5.47	113	259615	51.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.50%	
50) Toluene-d8	8.70	98	914422	48.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.92%	
62) 4-Bromofluorobenzene	11.12	95	337916	45.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.04%	
Target Compounds						
16) Acetone	2.42	43	52836	20.652	ug/l	93
18) Methyl Acetate	2.75	43	3188	0.516	ug/l #	88
52) Toluene	8.76	92	9269	0.725	ug/l	97
68) m/p-Xylenes	10.35	106	3216	0.355	ug/l	88
69) o-Xylene	10.68	106	2665	0.303	ug/l	75
84) 1,2,4-Trimethylbenzene	11.79	105	16004	0.841	ug/l	99
95) Naphthalene	13.82	128	10718	0.472	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX083120\
Data File : VX018175.D
Acq On : 01 Sep 2020 02:55
Operator : JC/SP
Sample : L3854-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1332-FM648

Quant Time: Sep 01 07:54:48 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

