

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083121\
 Data File : VX024030.D
 Acq On : 31 Aug 2021 09:34
 Operator : JC/MD
 Sample : VX0831MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0831MBL01

Quant Time: Sep 01 04:57:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	163251	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	277801	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	260033	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	106267	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	120836	53.034	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.060%
35) Dibromofluoromethane	5.391	113	92955	50.555	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.100%
50) Toluene-d8	8.653	98	341824	48.845	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.680%
62) 4-Bromofluorobenzene	11.085	95	123138	47.798	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083121\
Data File : VX024030.D
Acq On : 31 Aug 2021 09:34
Operator : JC/MD
Sample : VX0831MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0831MBL01

Quant Time: Sep 01 04:57:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

