

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083121\
 Data File : VX024037.D
 Acq On : 31 Aug 2021 12:27
 Operator : JC/MD
 Sample : M3577-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FIELD-BLANK

Quant Time: Sep 01 05:00:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	161021	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	273786	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	254997	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	110598	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	116851	51.995	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.000%
35) Dibromofluoromethane	5.397	113	92124	50.838	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.680%
50) Toluene-d8	8.653	98	336260	48.754	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.500%
62) 4-Bromofluorobenzene	11.085	95	123273	48.552	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.100%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	2427	2.415	ug/l	98
20) Methylene Chloride	2.788	84	1875	0.977	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083121\
Data File : VX024037.D
Acq On : 31 Aug 2021 12:27
Operator : JC/MD
Sample : M3577-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FIELD-BLANK

Quant Time: Sep 01 05:00:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

