

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083121\
 Data File : VX024040.D
 Acq On : 31 Aug 2021 13:38
 Operator : JC/MD
 Sample : M3584-14 40X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP02-20210827

Quant Time: Sep 01 05:01:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	143823	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	248345	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	231672	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96014	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	108887	54.245	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 108.500%		
35) Dibromofluoromethane	5.397	113	81617	49.653	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 99.300%		
50) Toluene-d8	8.653	98	302763	48.395	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 96.780%		
62) 4-Bromofluorobenzene	11.085	95	108803	47.243	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 94.480%		
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	2.325	96	775	0.561	ug/l	93
16) Acetone	2.386	43	2411	2.686	ug/l	91
20) Methylene Chloride	2.794	84	1486	0.867	ug/l	92
21) trans-1,2-Dichloroethene	3.105	96	856	0.581	ug/l #	70
27) cis-1,2-Dichloroethene	4.501	96	35063	19.987	ug/l	93
44) Trichloroethene	7.135	130	168587	89.740	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083121\
Data File : VX024040.D
Acq On : 31 Aug 2021 13:38
Operator : JC/MD
Sample : M3584-14 40X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP02-20210827

Quant Time: Sep 01 05:01:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

