

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083121\
 Data File : VX024045.D
 Acq On : 31 Aug 2021 15:36
 Operator : JC/MD
 Sample : M3584-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-140B-20210827

Quant Time: Sep 01 05:03:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	141735	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	242666	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	222594	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82514	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	105590	53.378	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.760%	
35) Dibromofluoromethane	5.404	113	81073	50.477	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.960%	
50) Toluene-d8	8.653	98	296145	48.444	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 96.880%	
62) 4-Bromofluorobenzene	11.085	95	100091	44.477	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 88.960%	
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	5373	6.075	ug/l #	86
27) cis-1,2-Dichloroethene	4.495	96	1518	0.878	ug/l	97
44) Trichloroethene	7.141	130	2271	1.237	ug/l	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083121\
Data File : VX024045.D
Acq On : 31 Aug 2021 15:36
Operator : JC/MD
Sample : M3584-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-140B-20210827

Quant Time: Sep 01 05:03:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

