

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083121\
 Data File : VX024046.D
 Acq On : 31 Aug 2021 16:00
 Operator : JC/MD
 Sample : M3584-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB01-20210827

Quant Time: Sep 01 05:03:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	145754	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	251088	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	232330	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96331	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	109075	53.619	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 107.240%			
35) Dibromofluoromethane	5.397	113	83872	50.468	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.940%			
50) Toluene-d8	8.653	98	307884	48.675	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 97.360%			
62) 4-Bromofluorobenzene	11.085	95	109433	46.997	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 94.000%			
Target Compounds						Qvalue
11) Tert butyl alcohol	2.977	59	2964	7.205	ug/l	95
16) Acetone	2.386	43	3136	3.448	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083121\
Data File : VX024046.D
Acq On : 31 Aug 2021 16:00
Operator : JC/MD
Sample : M3584-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB01-20210827

Quant Time: Sep 01 05:03:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

