

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083121\
 Data File : VX024047.D
 Acq On : 31 Aug 2021 16:24
 Operator : JC/MD
 Sample : M3584-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB01-20210827

Quant Time: Sep 01 05:04:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	141805	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	244419	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	226448	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85972	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	107573	54.353	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.700%	
35) Dibromofluoromethane	5.397	113	82160	50.787	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.580%	
50) Toluene-d8	8.653	98	300977	48.882	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 97.760%	
62) 4-Bromofluorobenzene	11.085	95	102996	45.440	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 90.880%	
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	3169	3.581	ug/l	97
20) Methylene Chloride	2.800	84	2277	1.347	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083121\
Data File : VX024047.D
Acq On : 31 Aug 2021 16:24
Operator : JC/MD
Sample : M3584-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB01-20210827

Quant Time: Sep 01 05:04:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

