

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083122\
 Data File : VX031015.D
 Acq On : 01 Sep 2022 06:46
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Sep 06 02:34:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 30 14:24:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	188574	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	313258	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	288352	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	149419	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	116179	42.955	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.900%
35) Dibromofluoromethane	5.391	113	109652	47.485	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.980%
50) Toluene-d8	8.653	98	394731	46.158	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.320%
62) 4-Bromofluorobenzene	11.085	95	139598	45.840	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.680%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	90812	45.520	ug/l	99
3) Chloromethane	1.294	50	75427	37.513	ug/l	97
4) Vinyl Chloride	1.374	62	94641	39.842	ug/l	99
5) Bromomethane	1.611	94	37498	51.156	ug/l	99
6) Chloroethane	1.685	64	38793	27.161	ug/l	99
7) Trichlorofluoromethane	1.886	101	163527	45.263	ug/l	97
8) Diethyl Ether	2.136	74	57915	41.538	ug/l	76
9) 1,1,2-Trichlorotrifluo...	2.331	101	105844	45.558	ug/l	95
10) Methyl Iodide	2.453	142	77498	37.157	ug/l	96
11) Tert butyl alcohol	3.001	59	144977	226.245	ug/l #	86
12) 1,1-Dichloroethene	2.319	96	104724	45.696	ug/l	98
13) Acrolein	2.239	56	66106	204.578	ug/l	99
14) Allyl chloride	2.666	41	134650	36.215	ug/l #	95
15) Acrylonitrile	3.069	53	286945	202.787	ug/l	98
16) Acetone	2.392	43	236769	193.960	ug/l	99
17) Carbon Disulfide	2.514	76	260439	43.059	ug/l	99
18) Methyl Acetate	2.709	43	136459	38.106	ug/l	91
19) Methyl tert-butyl Ether	3.117	73	336462	45.075	ug/l	95
20) Methylene Chloride	2.794	84	117922	42.488	ug/l	90
21) trans-1,2-Dichloroethene	3.093	96	110749	44.568	ug/l	94
22) Diisopropyl ether	3.764	45	296266	38.670	ug/l	95
23) Vinyl Acetate	3.727	43	1202835	195.698	ug/l #	92
24) 1,1-Dichloroethane	3.617	63	192690	42.947	ug/l	99
25) 2-Butanone	4.568	43	362910	183.646	ug/l #	87
26) 2,2-Dichloropropane	4.477	77	79552	23.529	ug/l	97
27) cis-1,2-Dichloroethene	4.495	96	133536	46.633	ug/l	91
28) Bromochloromethane	4.904	49	55791	38.442	ug/l #	71
29) Tetrahydrofuran	5.019	42	228301	178.585	ug/l #	83
30) Chloroform	5.099	83	210453	45.216	ug/l	98
31) Cyclohexane	5.477	56	159464	39.835	ug/l #	85
32) 1,1,1-Trichloroethane	5.385	97	182538	46.264	ug/l	96
36) 1,1-Dichloropropene	5.696	75	149917	43.680	ug/l	99
37) Ethyl Acetate	4.721	43	139934	38.782	ug/l	97
38) Carbon Tetrachloride	5.684	117	159900	48.937	ug/l	97
39) Methylcyclohexane	7.385	83	185180	44.111	ug/l	95
40) Benzene	6.044	78	452802	45.775	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083122\
 Data File : VX031015.D
 Acq On : 01 Sep 2022 06:46
 Operator : JC/MD
 Sample : VSTDC0050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDC0050EC

Quant Time: Sep 06 02:34:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 30 14:24:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	76241	38.903	ug/l	90
42) 1,2-Dichloroethane	6.092	62	148271	45.123	ug/l	100
43) Isopropyl Acetate	6.342	43	213246	38.086	ug/l #	92
44) Trichloroethene	7.129	130	138555	47.545	ug/l	99
45) 1,2-Dichloropropane	7.434	63	114012	43.180	ug/l	99
46) Dibromomethane	7.586	93	84670	46.957	ug/l	93
47) Bromodichloromethane	7.824	83	162279	46.814	ug/l	97
48) Methyl methacrylate	7.696	41	104667	37.347	ug/l #	85
49) 1,4-Dioxane	7.696	88	63793	866.123	ug/l	92
51) 4-Methyl-2-Pentanone	8.574	43	707038	191.324	ug/l	90
52) Toluene	8.720	92	291981	48.308	ug/l	98
53) t-1,3-Dichloropropene	8.982	75	153807	43.973	ug/l	98
54) cis-1,3-Dichloropropene	8.366	75	169999	42.484	ug/l #	91
55) 1,1,2-Trichloroethane	9.153	97	125805	47.985	ug/l	98
56) Ethyl methacrylate	9.116	69	180959	47.128	ug/l #	84
57) 1,3-Dichloropropane	9.311	76	194438	44.967	ug/l	97
58) 2-Chloroethyl Vinyl ether	8.244	63	424296	220.012	ug/l	94
59) 2-Hexanone	9.433	43	537989	195.773	ug/l	88
60) Dibromochloromethane	9.525	129	132061	50.683	ug/l	98
61) 1,2-Dibromoethane	9.610	107	134195	49.577	ug/l	99
64) Tetrachloroethene	9.275	164	136860	50.180	ug/l	96
65) Chlorobenzene	10.079	112	319456	46.853	ug/l	97
66) 1,1,1,2-Tetrachloroethane	10.165	131	120669	49.950	ug/l	99
67) Ethyl Benzene	10.195	91	571854	47.283	ug/l	98
68) m/p-Xylenes	10.305	106	444952	97.129	ug/l	99
69) o-Xylene	10.646	106	217623	48.539	ug/l	98
70) Styrene	10.659	104	366525	48.426	ug/l	99
71) Bromoform	10.799	173	99482	52.479	ug/l #	100
73) Isopropylbenzene	10.963	105	572823	48.458	ug/l	100
74) N-amyl acetate	10.848	43	183342	40.117	ug/l #	86
75) 1,1,2,2-Tetrachloroethane	11.213	83	176670	43.624	ug/l	99
76) 1,2,3-Trichloropropane	11.244	75	191158	51.966	ug/l	87
77) Bromobenzene	11.201	156	141298	49.326	ug/l	91
78) n-propylbenzene	11.305	91	656709	47.403	ug/l	99
79) 2-Chlorotoluene	11.366	91	390335	45.363	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	479003	48.089	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	44154	35.687	ug/l	93
82) 4-Chlorotoluene	11.457	91	439894	45.730	ug/l	99
83) tert-Butylbenzene	11.713	119	462447	47.295	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	481464	47.215	ug/l	98
85) sec-Butylbenzene	11.890	105	602630	47.484	ug/l	100
86) p-Isopropyltoluene	12.012	119	496699	47.176	ug/l	98
87) 1,3-Dichlorobenzene	11.969	146	266066	47.108	ug/l	98
88) 1,4-Dichlorobenzene	12.043	146	265479	45.843	ug/l	98
89) n-Butylbenzene	12.335	91	424823	43.111	ug/l	99
90) Hexachloroethane	12.542	117	82404	44.578	ug/l	92
91) 1,2-Dichlorobenzene	12.335	146	264860	46.230	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	38943	38.549	ug/l	88
93) 1,2,4-Trichlorobenzene	13.591	180	181517	49.562	ug/l	98
94) Hexachlorobutadiene	13.725	225	73981	47.569	ug/l	99
95) Naphthalene	13.780	128	683080	52.151	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	184266	49.832	ug/l	97


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083122\
Data File : VX031015.D
Acq On : 01 Sep 2022 06:46
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Quant Time: Sep 06 02:34:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 30 14:24:53 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:168 Resp: 188574
Ion Ratio Lower Upper
168 100
99 53.5 44.0 66.0



#2
Dichlorodifluoromethane
Concen: 45.520 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46

Tgt Ion: 85 Resp: 90812
Ion Ratio Lower Upper
85 100
87 33.4 16.5 49.5









#3
Chloromethane
Concen: 37.513 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.006 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 50 Resp: 75427
Ion Ratio Lower Upper
50 100
52 34.8 26.3 39.5



#4
Vinyl Chloride
Concen: 39.842 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46

Tgt Ion: 62 Resp: 94641
Ion Ratio Lower Upper
62 100
64 32.3 25.2 37.8









#5
Bromomethane
Concen: 51.156 ug/l
RT: 1.611 min Scan# 80
Delta R.T. 0.006 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 94 Resp: 37498
Ion Ratio Lower Upper
94 100
96 97.1 76.9 115.3



#6
Chloroethane
Concen: 27.161 ug/l
RT: 1.685 min Scan# 98
Delta R.T. -0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46

Tgt Ion: 64 Resp: 38793
Ion Ratio Lower Upper
64 100
66 33.2 27.2 40.8









#7
 Trichlorofluoromethane
 Concen: 45.263 ug/l
 RT: 1.886 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX031015.D
 Acq: 01 Sep 2022 06:46




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:101 Resp: 163527
 Ion Ratio Lower Upper
 101 100
 103 64.7 49.7 74.5



#8
 Diethyl Ether
 Concen: 41.538 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. -0.000 min
 Lab File: VX031015.D
 Acq: 01 Sep 2022 06:46

Tgt Ion: 74 Resp: 57915
 Ion Ratio Lower Upper
 74 100
 45 78.6 51.5 154.7









#9
1,1,2-Trichlorotrifluoroethane
Concen: 45.558 ug/l
RT: 2.331 min Scan# 203
Delta R.T. 0.006 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:101 Resp: 105844
Ion Ratio Lower Upper
101 100
85 41.0 34.0 51.0
151 80.0 59.4 89.0



#10
Methyl Iodide
Concen: 37.157 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46

Tgt Ion:142 Resp: 77498
Ion Ratio Lower Upper
142 100
127 40.9 30.2 45.2
141 13.7 11.3 16.9









#11
Tert butyl alcohol
Concen: 226.245 ug/l
RT: 3.001 min Scan# 311
Delta R.T. 0.012 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 59 Resp: 144977
Ion Ratio Lower Upper
59 100
57 4.1 7.4 11.2#



#12
1,1-Dichloroethene
Concen: 45.696 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46

Tgt Ion: 96 Resp: 104724
Ion Ratio Lower Upper
96 100
61 143.7 117.4 176.0
98 64.6 52.2 78.4









#13
Acrolein
Concen: 204.578 ug/l
RT: 2.239 min Scan# 189
Delta R.T. 0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 56 Resp: 66106
Ion Ratio Lower Upper
56 100
55 70.3 56.6 84.8



#14
Allyl chloride
Concen: 36.215 ug/l
RT: 2.666 min Scan# 259
Delta R.T. 0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46

Tgt Ion: 41 Resp: 134650
Ion Ratio Lower Upper
41 100
39 61.3 48.9 73.3
76 41.4 26.9 40.3#









#15
Acrylonitrile
Concen: 202.787 ug/l
RT: 3.069 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 53 Resp: 286945
Ion Ratio Lower Upper
53 100
52 83.1 65.4 98.0
51 35.6 27.7 41.5



#16
Acetone
Concen: 193.960 ug/l
RT: 2.392 min Scan# 214
Delta R.T. 0.006 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46

Tgt Ion: 43 Resp: 236769
Ion Ratio Lower Upper
43 100
58 35.2 27.7 41.5









#17
Carbon Disulfide
Concen: 43.059 ug/l
RT: 2.514 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 76 Resp: 260439
Ion Ratio Lower Upper
76 100
78 9.9 7.8 11.6



#18
Methyl Acetate
Concen: 38.106 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46

Tgt Ion: 43 Resp: 136459
Ion Ratio Lower Upper
43 100
74 28.8 19.4 29.2





#19

Methyl tert-butyl Ether

Concen: 45.075 ug/l

RT: 3.117 min Scan# 31

Delta R.T. 0.000 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 73 Resp: 336462

Ion Ratio Lower Upper

73 100

57 21.2 19.0 28.6



#20

Methylene Chloride

Concen: 42.488 ug/l

RT: 2.794 min Scan# 280

Delta R.T. 0.006 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46

Tgt Ion: 84 Resp: 117922




Ion Ratio Lower Upper


84 100

49 101.8 93.8 140.6

51 31.9 29.4 44.0

86 65.9 51.0 76.4









#21
trans-1,2-Dichloroethene
Concen: 44.568 ug/l
RT: 3.093 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 96 Resp: 110749
Ion Ratio Lower Upper
96 100
61 122.7 105.6 158.4
98 61.2 50.1 75.1



#22
Diisopropyl ether
Concen: 38.670 ug/l
RT: 3.764 min Scan# 439
Delta R.T. -0.006 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46

Tgt Ion: 45 Resp: 296266
Ion Ratio Lower Upper
45 100
43 78.4 65.4 98.0
87 33.4 22.3 33.5
59 13.0 9.0 13.4









#23
 Vinyl Acetate
 Concen: 195.698 ug/l
 RT: 3.727 min Scan# 411
 Delta R.T. -0.000 min
 Lab File: VX031015.D
 Acq: 01 Sep 2022 06:46




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 43 Resp: 1202835
 Ion Ratio Lower Upper
 43 100
 86 13.7 8.5 12.7#



#24
 1,1-Dichloroethane
 Concen: 42.947 ug/l
 RT: 3.617 min Scan# 415
 Delta R.T. 0.006 min
 Lab File: VX031015.D
 Acq: 01 Sep 2022 06:46

Tgt Ion: 63 Resp: 192690
 Ion Ratio Lower Upper
 63 100
 98 7.5 3.5 10.5
 100 4.7 2.3 6.9





#25

2-Butanone

Concen: 183.646 ug/l

RT: 4.568 min Scan# 571

Delta R.T. 0.000 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 43 Resp: 362910

Ion Ratio Lower Upper

43 100

72 33.6 21.7 32.5#



#26

2,2-Dichloropropane

Concen: 23.529 ug/l

RT: 4.477 min Scan# 556

Delta R.T. -0.000 min

Lab File: VX031015.D




Acq: 01 Sep 2022 06:46


Tgt Ion: 77 Resp: 79552

Ion Ratio Lower Upper

77 100

97 27.7 13.0 39.0









#27
 cis-1,2-Dichloroethene
 Concen: 46.633 ug/l
 RT: 4.495 min Scan# 51
 Delta R.T. 0.006 min
 Lab File: VX031015.D
 Acq: 01 Sep 2022 06:46




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 96 Resp: 133536
 Ion Ratio Lower Upper
 96 100
 61 120.1 0.0 270.0
 98 64.3 0.0 127.8



#28
 Bromochloromethane
 Concen: 38.442 ug/l
 RT: 4.904 min Scan# 626
 Delta R.T. -0.000 min
 Lab File: VX031015.D
 Acq: 01 Sep 2022 06:46

Tgt Ion: 49 Resp: 55791
 Ion Ratio Lower Upper
 49 100
 129 2.4 0.0 4.4
 130 104.8 62.8 94.2#





#29

Tetrahydrofuran

Concen: 178.585 ug/l

RT: 5.019 min Scan# 644

Delta R.T. 0.006 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 42 Resp: 228301

Ion Ratio Lower Upper

42 100

72 57.3 36.5 54.7#

71 52.7 33.9 50.9#



#30

Chloroform

Concen: 45.216 ug/l

RT: 5.099 min Scan# 658

Delta R.T. -0.000 min

Lab File: VX031015.D




Acq: 01 Sep 2022 06:46


Tgt Ion: 83 Resp: 210453

Ion Ratio Lower Upper

83 100

85 66.3 51.5 77.3









#31
Cyclohexane
Concen: 39.835 ug/l
RT: 5.477 min Scan# 712
Delta R.T. -0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 56 Resp: 159464
Ion Ratio Lower Upper
56 100
69 34.2 25.4 38.0
84 103.4 68.9 103.3#



#32
1,1,1-Trichloroethane
Concen: 46.264 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.006 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46

Tgt Ion: 97 Resp: 182538
Ion Ratio Lower Upper
97 100
99 64.1 51.4 77.0
61 39.2 35.8 53.8





#33

1,2-Dichloroethane-d4

Concen: 42.955 ug/l

RT: 5.958 min Scan# 799

Delta R.T. 0.000 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 65 Resp: 116179

Ion Ratio Lower Upper

65 100

67 53.5 0.0 109.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.006 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46




Tgt Ion: 114 Resp: 313258


Ion Ratio Lower Upper

114 100

63 17.9 0.0 38.4

88 14.4 0.0 31.8









#35
Dibromofluoromethane
Concen: 47.485 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:113 Resp: 109652
Ion Ratio Lower Upper
113 100
111 101.2 83.2 124.8
192 19.3 14.8 22.2



#36
1,1-Dichloropropene
Concen: 43.680 ug/l
RT: 5.696 min Scan# 756
Delta R.T. 0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46

Tgt Ion: 75 Resp: 149917
Ion Ratio Lower Upper
75 100
110 39.2 19.3 57.9
77 30.9 24.4 36.6





#37

Ethyl Acetate

Concen: 38.782 ug/l

RT: 4.721 min Scan# 596

Delta R.T. -0.000 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 43 Resp: 139934

Ion Ratio Lower Upper

43 100

61 16.1 11.7 17.5

70 12.7 9.3 13.9



#38

Carbon Tetrachloride

Concen: 48.937 ug/l

RT: 5.684 min Scan# 754

Delta R.T. -0.000 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46




Tgt Ion: 117 Resp: 159900


Ion Ratio Lower Upper

117 100

119 95.7 74.1 111.1

121 29.9 25.8 38.8





#39

Methylcyclohexane

Concen: 44.111 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 83 Resp: 185180

Ion Ratio Lower Upper

83 100

55 65.0 57.7 86.5

98 45.6 36.5 54.7



#40

Benzene

Concen: 45.775 ug/l

RT: 6.044 min Scan# 813

Delta R.T. -0.000 min

Lab File: VX031015.D




Acq: 01 Sep 2022 06:46


Tgt Ion: 78 Resp: 452802

Ion Ratio Lower Upper

78 100

77 23.3 18.0 27.0









#41
Methacrylonitrile
Concen: 38.903 ug/l
RT: 4.922 min Scan# 61
Delta R.T. -0.006 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:	41	Resp:	76241
Ion	Ratio	Lower	Upper
41	100		
39	52.2	42.9	64.3
67	89.7	60.1	90.1
52	32.8	22.6	34.0



#42
1,2-Dichloroethane
Concen: 45.123 ug/l
RT: 6.092 min Scan# 821
Delta R.T. 0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46

Tgt Ion:	62	Resp:	148271
Ion	Ratio	Lower	Upper
62	100		
98	10.1	0.0	20.2









#43
Isopropyl Acetate
Concen: 38.086 ug/l
RT: 6.342 min Scan# 862
Delta R.T. -0.006 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 213246
Ion Ratio Lower Upper
43 100
61 24.2 16.5 24.7
87 17.3 11.1 16.7#



#44
Trichloroethene
Concen: 47.545 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46

Tgt Ion:130 Resp: 138555
Ion Ratio Lower Upper
130 100
95 96.1 0.0 194.8





#45

1,2-Dichloropropane

Concen: 43.180 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. -0.006 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 63 Resp: 114012

Ion Ratio Lower Upper

63 100

65 30.8 25.2 37.8



#46

Dibromomethane

Concen: 46.957 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. 0.000 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46




Tgt Ion: 93 Resp: 84670


Ion Ratio Lower Upper

93 100

95 85.1 67.4 101.0

174 105.8 75.6 113.4









#47
Bromodichloromethane
Concen: 46.814 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 83 Resp: 162279
Ion Ratio Lower Upper
83 100
85 63.9 53.5 80.3
127 8.9 6.6 10.0



#48
Methyl methacrylate
Concen: 37.347 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46

Tgt Ion: 41 Resp: 104667
Ion Ratio Lower Upper
41 100
69 106.6 68.8 103.2#
39 54.1 44.6 66.8





#49

1,4-Dioxane

Concen: 866.123 ug/l

RT: 7.696 min Scan# 1084

Delta R.T. 0.019 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 88 Resp: 63793

Ion Ratio Lower Upper

88 100

43 24.9 23.8 35.8

58 61.1 53.8 80.6



#50

Toluene-d8

Concen: 46.158 ug/l

RT: 8.653 min Scan# 1241

Delta R.T. -0.000 min

Lab File: VX031015.D




Acq: 01 Sep 2022 06:46


Tgt Ion: 98 Resp: 394731

Ion Ratio Lower Upper

98 100

100 64.9 52.3 78.5





#51

4-Methyl-2-Pentanone

Concen: 191.324 ug/l

RT: 8.574 min Scan# 1228

Delta R.T. -0.006 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 43 Resp: 707038

Ion Ratio Lower Upper

43 100

58 45.7 31.4 47.2



#52

Toluene

Concen: 48.308 ug/l

RT: 8.720 min Scan# 1252

Delta R.T. -0.000 min

Lab File: VX031015.D




Acq: 01 Sep 2022 06:46


Tgt Ion: 92 Resp: 291981

Ion Ratio Lower Upper

92 100

91 176.3 139.2 208.8





#53

t-1,3-Dichloropropene

Concen: 43.973 ug/l

RT: 8.982 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 75 Resp: 153807

Ion Ratio Lower Upper

75 100

77 31.6 26.3 39.5



#54

cis-1,3-Dichloropropene

Concen: 42.484 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.006 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46




Tgt Ion: 75 Resp: 169999


Ion Ratio Lower Upper

75 100

77 31.4 25.0 37.4

39 36.9 37.5 56.3#





#55

1,1,2-Trichloroethane

Concen: 47.985 ug/l

RT: 9.153 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 97 Resp: 125805





Ion Ratio Lower Upper

97 100

83 85.3 68.9 103.3

85 52.8 44.0 66.0

99 62.9 48.3 72.5



#56

Ethyl methacrylate

Concen: 47.128 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.006 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46




Tgt Ion: 69 Resp: 180959


Ion Ratio Lower Upper

69 100

41 52.3 52.3 78.5#

39 27.2 28.2 42.4#





#57

1,3-Dichloropropane

Concen: 44.967 ug/l

RT: 9.311 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 76 Resp: 194438

Ion Ratio Lower Upper

76 100

78 33.8 25.9 38.9



#58

2-Chloroethyl Vinyl ether

Concen: 220.012 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. -0.013 min

Lab File: VX031015.D




Acq: 01 Sep 2022 06:46


Tgt Ion: 63 Resp: 424296

Ion Ratio Lower Upper

63 100

106 30.7 22.2 33.2





#59

2-Hexanone

Concen: 195.773 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.000 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 43 Resp: 537989

Ion Ratio Lower Upper

43 100

58 63.8 27.6 82.8



#60

Dibromochloromethane

Concen: 50.683 ug/l

RT: 9.525 min Scan# 1384

Delta R.T. -0.000 min

Lab File: VX031015.D




Acq: 01 Sep 2022 06:46


Tgt Ion: 129 Resp: 132061

Ion Ratio Lower Upper

129 100

127 78.9 38.5 115.5





#61

1,2-Dibromoethane

Concen: 49.577 ug/l

RT: 9.610 min Scan# 1399

Delta R.T. 0.000 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion:107 Resp: 134195

Ion Ratio Lower Upper

107 100

109 94.4 76.6 114.8



#62

4-Bromofluorobenzene

Concen: 45.840 ug/l

RT: 11.085 min Scan# 1640

Delta R.T. 0.000 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46




Tgt Ion: 95 Resp: 139598


Ion Ratio Lower Upper

95 100

174 84.9 0.0 149.0

176 81.8 0.0 146.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:117 Resp: 288352
Ion Ratio Lower Upper
117 100
82 52.7 42.7 64.1
119 32.1 25.4 38.2



#64
Tetrachloroethene
Concen: 50.180 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46

Tgt Ion:164 Resp: 136860
Ion Ratio Lower Upper
164 100
166 124.2 98.2 147.4
129 92.9 80.8 121.2
131 90.9 75.5 113.3









#65
Chlorobenzene
Concen: 46.853 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:112 Resp: 319456
Ion Ratio Lower Upper
112 100
114 32.3 24.6 36.8



#66
1,1,1,2-Tetrachloroethane
Concen: 49.950 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46

Tgt Ion:131 Resp: 120669
Ion Ratio Lower Upper
131 100
133 94.0 46.9 140.8
119 67.0 34.4 103.1









#67
Ethyl Benzene
Concen: 47.283 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 571854
Ion Ratio Lower Upper
91 100
106 31.0 25.5 38.3



#68
m/p-Xylenes
Concen: 97.129 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46

Tgt Ion: 106 Resp: 444952
Ion Ratio Lower Upper
106 100
91 199.2 158.2 237.4





#69

o-Xylene

Concen: 48.539 ug/l

RT: 10.646 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion:106 Resp: 217623

Ion Ratio Lower Upper

106 100

91 211.4 103.8 311.6



#70

Styrene

Concen: 48.426 ug/l

RT: 10.659 min Scan# 1570

Delta R.T. -0.000 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46




Tgt Ion:104 Resp: 366525


Ion Ratio Lower Upper

104 100

78 46.3 37.8 56.8

103 55.5 44.5 66.7









#71
Bromoform
Concen: 52.479 ug/l
RT: 10.799 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:173 Resp: 99482
Ion Ratio Lower Upper
173 100
175 47.6 23.8 71.2
254 0.0 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46

Tgt Ion:152 Resp: 149419
Ion Ratio Lower Upper
152 100
115 81.5 40.5 121.5
150 170.6 0.0 347.0









#73
Isopropylbenzene
Concen: 48.458 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. 0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:105 Resp: 572823
Ion Ratio Lower Upper
105 100
120 26.7 13.5 40.4



#74
N-ethyl acetate
Concen: 40.117 ug/l
RT: 10.848 min Scan# 1601
Delta R.T. -0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46

Tgt Ion: 43 Resp: 183342
Ion Ratio Lower Upper
43 100
70 52.7 32.3 48.5#
55 27.0 18.4 27.6
61 28.3 19.2 28.8









#75
 1,1,2,2-Tetrachloroethane
 Concen: 43.624 ug/l
 RT: 11.213 min Scan# 1661
 Delta R.T. 0.000 min
 Lab File: VX031015.D
 Acq: 01 Sep 2022 06:46




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 83 Resp: 176670
 Ion Ratio Lower Upper
 83 100
 131 10.3 4.5 13.7
 85 64.3 32.5 97.5



#76
 1,2,3-Trichloropropane
 Concen: 51.966 ug/l
 RT: 11.244 min Scan# 1666
 Delta R.T. -0.000 min
 Lab File: VX031015.D
 Acq: 01 Sep 2022 06:46

Tgt Ion: 75 Resp: 191158
 Ion Ratio Lower Upper
 75 100
 77 31.7 19.8 59.3





#77

Bromobenzene

Concen: 49.326 ug/l

RT: 11.201 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:156 Resp: 141298

Ion Ratio Lower Upper

156 100

77 148.6 82.7 247.9

158 97.7 47.5 142.5



#78

n-propylbenzene

Concen: 47.403 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. -0.000 min

Lab File: VX031015.D




Acq: 01 Sep 2022 06:46


Tgt Ion: 91 Resp: 656709

Ion Ratio Lower Upper

91 100

120 23.8 11.6 34.8





#79

2-Chlorotoluene

Concen: 45.363 ug/l

RT: 11.366 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 91 Resp: 390335

Ion Ratio Lower Upper

91 100

126 35.4 17.4 52.3



#80

1,3,5-Trimethylbenzene

Concen: 48.089 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. -0.006 min

Lab File: VX031015.D




Acq: 01 Sep 2022 06:46


Tgt Ion: 105 Resp: 479003

Ion Ratio Lower Upper

105 100

120 49.0 24.4 73.2









#81
trans-1,4-Dichloro-2-butene
Concen: 35.687 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.006 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 75 Resp: 44154
Ion Ratio Lower Upper
75 100
53 86.1 74.2 111.4
89 47.9 34.8 52.2



#82
4-Chlorotoluene
Concen: 45.730 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46

Tgt Ion: 91 Resp: 439894
Ion Ratio Lower Upper
91 100
126 31.0 15.3 45.9









#83
 tert-Butylbenzene
 Concen: 47.295 ug/l
 RT: 11.713 min Scan# 11
 Delta R.T. -0.006 min
 Lab File: VX031015.D
 Acq: 01 Sep 2022 06:46




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:119 Resp: 462447
 Ion Ratio Lower Upper
 119 100
 91 57.2 28.3 84.9
 134 23.0 11.2 33.5



#84
 1,2,4-Trimethylbenzene
 Concen: 47.215 ug/l
 RT: 11.756 min Scan# 1750
 Delta R.T. 0.000 min
 Lab File: VX031015.D
 Acq: 01 Sep 2022 06:46

Tgt Ion:105 Resp: 481464
 Ion Ratio Lower Upper
 105 100
 120 44.7 22.9 68.8









#85
 sec-Butylbenzene
 Concen: 47.484 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.006 min
 Lab File: VX031015.D
 Acq: 01 Sep 2022 06:46




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:105 Resp: 602630
 Ion Ratio Lower Upper
 105 100
 134 19.9 10.0 30.0



#86
 p-Isopropyltoluene
 Concen: 47.176 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX031015.D
 Acq: 01 Sep 2022 06:46

Tgt Ion:119 Resp: 496699
 Ion Ratio Lower Upper
 119 100
 134 26.7 13.9 41.6
 91 22.7 11.8 35.3





#87

1,3-Dichlorobenzene

Concen: 47.108 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:146 Resp: 266066

Ion Ratio Lower Upper

146 100

111 38.6 20.3 60.8

148 64.1 31.8 95.3



#88

1,4-Dichlorobenzene

Concen: 45.843 ug/l

RT: 12.043 min Scan# 1797

Delta R.T. -0.000 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46




Tgt Ion:146 Resp: 265479


Ion Ratio Lower Upper

146 100

111 38.1 20.3 60.9

148 64.1 32.3 96.8





#89

n-Butylbenzene

Concen: 43.111 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 91 Resp: 424823

Ion Ratio Lower Upper

91 100

92 52.9 26.2 78.5

134 26.3 12.6 37.8



#90

Hexachloroethane

Concen: 44.578 ug/l

RT: 12.542 min Scan# 1879

Delta R.T. 0.000 min

Lab File: VX031015.D




Acq: 01 Sep 2022 06:46


Tgt Ion: 117 Resp: 82404

Ion Ratio Lower Upper

117 100

201 75.0 34.2 102.5









#91
1,2-Dichlorobenzene
Concen: 46.230 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:146 Resp: 264860
Ion Ratio Lower Upper
146 100
111 40.3 20.7 62.1
148 64.0 32.0 96.2



#92
1,2-Dibromo-3-Chloropropane
Concen: 38.549 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46

Tgt Ion: 75 Resp: 38943
Ion Ratio Lower Upper
75 100
155 101.8 45.0 134.8
157 126.4 56.9 170.7





#93

1,2,4-Trichlorobenzene

Concen: 49.562 ug/l

RT: 13.591 min Scan# 2051

Delta R.T. 0.000 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:180 Resp: 181517

Ion Ratio Lower Upper

180 100

182 98.3 48.2 144.6

145 31.6 15.9 47.6



#94

Hexachlorobutadiene

Concen: 47.569 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.000 min

Lab File: VX031015.D

Acq: 01 Sep 2022 06:46




Tgt Ion:225 Resp: 73981


Ion Ratio Lower Upper

225 100

223 62.6 30.6 91.6

227 63.1 31.9 95.9









#95
Naphthalene
Concen: 52.151 ug/l
RT: 13.780 min Scan# 2081
Delta R.T. 0.006 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion:128 Resp: 683080
Ion Ratio Lower Upper
128 100
127 12.5 10.2 15.2
129 10.8 8.6 13.0



#96
1,2,3-Trichlorobenzene
Concen: 49.832 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX031015.D
Acq: 01 Sep 2022 06:46

Tgt Ion:180 Resp: 184266
Ion Ratio Lower Upper
180 100
182 97.6 47.0 141.0
145 33.9 17.0 50.9

