

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083122\
 Data File : VX030976.D
 Acq On : 31 Aug 2022 14:23
 Operator : JC/MD
 Sample : N4381-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-38

Quant Time: Sep 01 00:56:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 30 14:24:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	209271	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	349386	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	310864	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	141686	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	145489	48.471	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	96.940%		
35) Dibromofluoromethane	5.391	113	135969	52.794	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	105.580%		
50) Toluene-d8	8.652	98	498157	52.229	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	104.460%		
62) 4-Bromofluorobenzene	11.085	95	164863	48.539	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	97.080%		
Target Compounds						
16) Acetone	2.385	43	3634	2.683	ug/l #	87
19) Methyl tert-butyl Ether	3.123	73	2395	0.289	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083122\
Data File : VX030976.D
Acq On : 31 Aug 2022 14:23
Operator : JC/MD
Sample : N4381-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-38

Quant Time: Sep 01 00:56:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 30 14:24:53 2022
Response via : Initial Calibration

