

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX083122\  
 Data File : VX030978.D  
 Acq On : 31 Aug 2022 15:10  
 Operator : JC/MD  
 Sample : N4381-09  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 14 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 MW-42

Quant Time: Sep 01 00:57:03 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X083022W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Aug 30 14:24:53 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min)    |
|-----------------------------|--------|-------|----------|----------|-------|-------------|
| -----                       |        |       |          |          |       |             |
| Internal Standards          |        |       |          |          |       |             |
| 1) Pentafluorobenzene       | 5.562  | 168   | 197946   | 50.000   | ug/l  | 0.00        |
| 34) 1,4-Difluorobenzene     | 6.769  | 114   | 343020   | 50.000   | ug/l  | 0.00        |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 302004   | 50.000   | ug/l  | 0.00        |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 137286   | 50.000   | ug/l  | 0.00        |
| System Monitoring Compounds |        |       |          |          |       |             |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 141680   | 49.903   | ug/l  | 0.00        |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | =     | 99.800%     |
| 35) Dibromofluoromethane    | 5.391  | 113   | 125513   | 49.638   | ug/l  | 0.00        |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =     | 99.280%     |
| 50) Toluene-d8              | 8.653  | 98    | 473403   | 50.555   | ug/l  | 0.00        |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | =     | 101.100%    |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 159697   | 47.890   | ug/l  | 0.00        |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =     | 95.780%     |
| Target Compounds            |        |       |          |          |       |             |
| 17) Carbon Disulfide        | 2.514  | 76    | 2004     | 0.316    | ug/l  | Qvalue # 92 |
| -----                       |        |       |          |          |       |             |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX083122\  
Data File : VX030978.D  
Acq On : 31 Aug 2022 15:10  
Operator : JC/MD  
Sample : N4381-09  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 14 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
MW-42

Quant Time: Sep 01 00:57:03 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X083022W.M  
Quant Title : SW846 8260  
QLast Update : Tue Aug 30 14:24:53 2022  
Response via : Initial Calibration

