

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083122\
 Data File : VX030985.D
 Acq On : 31 Aug 2022 17:54
 Operator : JC/MD
 Sample : N4381-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-50

Quant Time: Sep 01 00:58:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 30 14:24:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	198395	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	336773	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	297854	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	132389	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	132627	46.609	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.220%
35) Dibromofluoromethane	5.391	113	121890	49.100	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.200%
50) Toluene-d8	8.653	98	453856	49.366	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.740%
62) 4-Bromofluorobenzene	11.085	95	151035	46.133	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.260%
Target Compounds						
					Qvalue	
3) Chloromethane	1.288	50	1116	0.528	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083122\
Data File : VX030985.D
Acq On : 31 Aug 2022 17:54
Operator : JC/MD
Sample : N4381-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-50

Quant Time: Sep 01 00:58:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 30 14:24:53 2022
Response via : Initial Calibration

