

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083122\
 Data File : VX030990.D
 Acq On : 31 Aug 2022 19:52
 Operator : JC/MD
 Sample : N4381-08DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-34DL

Quant Time: Sep 01 00:58:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 30 14:24:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	205253	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	348758	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	304011	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	138118	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	136122	46.238	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.480%
35) Dibromofluoromethane	5.397	113	125612	48.860	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.720%
50) Toluene-d8	8.653	98	461300	48.452	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.900%
62) 4-Bromofluorobenzene	11.085	95	152403	44.951	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.900%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.971	59	38224	54.804	ug/l #	74
19) Methyl tert-butyl Ether	3.117	73	701598	86.354	ug/l	95
20) Methylene Chloride	2.788	84	1307	0.433	ug/l	90
40) Benzene	6.050	78	13326	1.210	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083122\
Data File : VX030990.D
Acq On : 31 Aug 2022 19:52
Operator : JC/MD
Sample : N4381-08DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-34DL

Quant Time: Sep 01 00:58:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 30 14:24:53 2022
Response via : Initial Calibration

