

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083122\
 Data File : VX030994.D
 Acq On : 31 Aug 2022 22:36
 Operator : JC/MD
 Sample : VX0831WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0831WBL02

Quant Time: Sep 01 00:59:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 30 14:24:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	217577	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	375963	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	326041	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	140030	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	141720	45.413	ug/l	0.01
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.820%
35) Dibromofluoromethane	5.391	113	131776	47.549	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.100%
50) Toluene-d8	8.653	98	494257	48.157	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.320%
62) 4-Bromofluorobenzene	11.085	95	160353	43.874	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083122\
Data File : VX030994.D
Acq On : 31 Aug 2022 22:36
Operator : JC/MD
Sample : VX0831WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0831WBL02

Quant Time: Sep 01 00:59:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 30 14:24:53 2022
Response via : Initial Calibration

