

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083122\
 Data File : VX031002.D
 Acq On : 01 Sep 2022 01:43
 Operator : JC/MD
 Sample : N4380-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-8

Quant Time: Sep 01 02:24:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 30 14:24:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	228887	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.769	114	369877	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	335861	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	146632	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	148176	45.136	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.280%
35) Dibromofluoromethane	5.391	113	134546	49.347	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.700%
50) Toluene-d8	8.653	98	498721	49.391	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.780%
62) 4-Bromofluorobenzene	11.085	95	168344	46.818	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083122\
Data File : VX031002.D
Acq On : 01 Sep 2022 01:43
Operator : JC/MD
Sample : N4380-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-8

Quant Time: Sep 01 02:24:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 30 14:24:53 2022
Response via : Initial Calibration

