

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083122\
 Data File : VX031003.D
 Acq On : 01 Sep 2022 02:06
 Operator : JC/MD
 Sample : N4380-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-27

Quant Time: Sep 01 04:30:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 30 14:24:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	222575	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	374094	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	323164	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	135941	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	140004	43.856	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.720%
35) Dibromofluoromethane	5.391	113	131415	47.655	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.320%
50) Toluene-d8	8.653	98	492077	48.184	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.360%
62) 4-Bromofluorobenzene	11.085	95	157937	43.428	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.860%
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.983	59	44099	58.306	ug/l	97
19) Methyl tert-butyl Ether	3.117	73	397814	45.153	ug/l	93
40) Benzene	6.050	78	4943	0.418	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083122\
Data File : VX031003.D
Acq On : 01 Sep 2022 02:06
Operator : JC/MD
Sample : N4380-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-27

Quant Time: Sep 01 04:30:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 30 14:24:53 2022
Response via : Initial Calibration

