

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083122\
 Data File : VX031005.D
 Acq On : 01 Sep 2022 02:53
 Operator : JC/MD
 Sample : N4380-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-41

Quant Time: Sep 01 04:30:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 30 14:24:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	225223	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	380982	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	344208	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	150602	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	143551	44.438	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	88.880%		
35) Dibromofluoromethane	5.391	113	134620	47.935	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.860%		
50) Toluene-d8	8.653	98	503127	48.375	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	96.760%		
62) 4-Bromofluorobenzene	11.085	95	171399	46.278	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	92.560%		
Target Compounds						
16) Acetone	2.392	43	2893	1.984	ug/l	97
19) Methyl tert-butyl Ether	3.111	73	6158	0.691	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083122\
Data File : VX031005.D
Acq On : 01 Sep 2022 02:53
Operator : JC/MD
Sample : N4380-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-41

Quant Time: Sep 01 04:30:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 30 14:24:53 2022
Response via : Initial Calibration

