

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083122\
 Data File : VX031007.D
 Acq On : 01 Sep 2022 03:40
 Operator : JC/MD
 Sample : N4454-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-13A-8-13-083022

Quant Time: Sep 01 04:30:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 30 14:24:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	228847	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	380940	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	340803	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	150992	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	141445	43.093	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	86.180%		
35) Dibromofluoromethane	5.385	113	134394	47.860	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.720%		
50) Toluene-d8	8.653	98	497023	47.794	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	95.580%		
62) 4-Bromofluorobenzene	11.085	95	167796	45.310	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	90.620%		
Target Compounds						Qvalue
4) Vinyl Chloride	1.374	62	88730	30.780	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.325	101	867	0.308	ug/l #	90
12) 1,1-Dichloroethene	2.325	96	4487	1.613	ug/l	84
17) Carbon Disulfide	2.532	76	6308	0.859	ug/l #	90
21) trans-1,2-Dichloroethene	3.099	96	4082	1.354	ug/l	94
24) 1,1-Dichloroethane	3.617	63	4345	0.798	ug/l #	93
27) cis-1,2-Dichloroethene	4.495	96	400971	115.383	ug/l	87
40) Benzene	6.050	78	9041	0.752	ug/l	98
44) Trichloroethene	7.129	130	143298	40.436	ug/l	93
64) Tetrachloroethene	9.281	164	1084	0.336	ug/l #	67
65) Chlorobenzene	10.079	112	15772	1.957	ug/l	96
88) 1,4-Dichlorobenzene	12.042	146	6686	1.143	ug/l #	79
91) 1,2-Dichlorobenzene	12.335	146	2389	0.413	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083122\
Data File : VX031007.D
Acq On : 01 Sep 2022 03:40
Operator : JC/MD
Sample : N4454-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-13A-8-13-083022

Quant Time: Sep 01 04:30:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 30 14:24:53 2022
Response via : Initial Calibration

