

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083122\
 Data File : VX031008.D
 Acq On : 01 Sep 2022 04:03
 Operator : JC/MD
 Sample : N4454-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OW-3B-49-54-083022

Quant Time: Sep 01 04:30:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 30 14:24:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	223268	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	367210	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	336833	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	146072	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	143222	44.725	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	89.440%		
35) Dibromofluoromethane	5.391	113	133881	49.460	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.920%		
50) Toluene-d8	8.653	98	494798	49.359	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.720%		
62) 4-Bromofluorobenzene	11.085	95	165194	46.275	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	92.560%		
Target Compounds						
27) cis-1,2-Dichloroethene	4.501	96	949	0.280	ug/l	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083122\
Data File : VX031008.D
Acq On : 01 Sep 2022 04:03
Operator : JC/MD
Sample : N4454-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OW-3B-49-54-083022

Quant Time: Sep 01 04:30:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 30 14:24:53 2022
Response via : Initial Calibration

