

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083122\
 Data File : VX031011.D
 Acq On : 01 Sep 2022 05:13
 Operator : JC/MD
 Sample : N4454-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-07-1.5-11.5-082922

Quant Time: Sep 01 06:16:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 30 14:24:53 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 09/01/2022
 Supervised By :Mahesh Dadoda 09/01/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	216997	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	358665	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	319776	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	139424	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	138865	44.617	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	89.240%		
35) Dibromofluoromethane	5.391	113	131157	49.608	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.220%		
50) Toluene-d8	8.653	98	479103	48.932	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	97.860%		
62) 4-Bromofluorobenzene	11.079	95	162665	46.653	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	93.300%		
Target Compounds						
4) Vinyl Chloride	1.368	62	1383	0.506	ug/l #	70
27) cis-1,2-Dichloroethene	4.489	96	1381m	0.419	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083122\
Data File : VX031011.D
Acq On : 01 Sep 2022 05:13
Operator : JC/MD
Sample : N4454-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-07-1.5-11.5-082922

Quant Time: Sep 01 06:16:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 30 14:24:53 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 09/01/2022
Supervised By :Mahesh Dadoda 09/01/2022

