

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX083122\
 Data File : VX030966.D
 Acq On : 31 Aug 2022 09:13
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 01 00:54:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 30 14:24:53 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	99	0.00
2 T	Dichlorodifluoromethane	50.000	49.991	0.0	98	0.00
3 P	Chloromethane	50.000	44.202	11.6	93	0.00
4 C	Vinyl Chloride	50.000	45.590	8.8#	96	0.00
5 T	Bromomethane	50.000	61.757	-23.5	118	0.00
6 T	Chloroethane	50.000	34.558	30.9#	75	0.00
7 T	Trichlorofluoromethane	50.000	47.263	5.5	100	0.00
8 T	Diethyl Ether	50.000	44.386	11.2	96	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.976	0.0	107	0.00
10 T	Methyl Iodide	50.000	54.989	-10.0	115	0.00
11 T	Tert butyl alcohol	250.000	199.853	20.1	84	0.00
12 CM	1,1-Dichloroethene	50.000	48.621	2.8#	103	0.00
13 T	Acrolein	250.000	252.955	-1.2	106	0.00
14 T	Allyl chloride	50.000	46.316	7.4	95	0.00
15 T	Acrylonitrile	250.000	224.869	10.1	91	0.00
16 T	Acetone	250.000	228.298	8.7	98	0.00
17 T	Carbon Disulfide	50.000	47.428	5.1	101	0.00
18 T	Methyl Acetate	50.000	42.588	14.8	91	0.00
19 T	Methyl tert-butyl Ether	50.000	47.746	4.5	97	0.00
20 T	Methylene Chloride	50.000	44.778	10.4	100	0.00
21 T	trans-1,2-Dichloroethene	50.000	48.028	3.9	101	0.00
22 T	Diisopropyl ether	50.000	44.702	10.6	94	0.00
23 T	Vinyl Acetate	250.000	229.561	8.2	92	0.00
24 P	1,1-Dichloroethane	50.000	46.880	6.2	97	0.00
25 T	2-Butanone	250.000	219.642	12.1	90	0.00
26 T	2,2-Dichloropropane	50.000	48.893	2.2	102	0.00
27 T	cis-1,2-Dichloroethene	50.000	48.658	2.7	100	0.00
28 T	Bromochloromethane	50.000	47.671	4.7	106	0.00
29 T	Tetrahydrofuran	250.000	209.468	16.2	86	0.00
30 C	Chloroform	50.000	47.154	5.7#	100	0.00
31 T	Cyclohexane	50.000	46.250	7.5	96	0.00
32 T	1,1,1-Trichloroethane	50.000	48.276	3.4	99	0.00
33 S	1,2-Dichloroethane-d4	50.000	46.366	7.3	98	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	97	0.00
35 S	Dibromofluoromethane	50.000	49.587	0.8	102	0.00
36 T	1,1-Dichloropropene	50.000	48.273	3.5	100	0.00
37 T	Ethyl Acetate	50.000	45.070	9.9	90	0.00
38 T	Carbon Tetrachloride	50.000	51.062	-2.1	100	0.00
39 T	Methylcyclohexane	50.000	51.655	-3.3	102	0.00
40 TM	Benzene	50.000	49.475	1.0	100	0.00
41 T	Methacrylonitrile	50.000	45.217	9.6	89	0.00
42 TM	1,2-Dichloroethane	50.000	48.737	2.5	97	0.00
43 T	Isopropyl Acetate	50.000	45.520	9.0	91	-0.01
44 TM	Trichloroethene	50.000	50.369	-0.7	106	0.00
45 C	1,2-Dichloropropane	50.000	48.516	3.0#	98	0.00
46 T	Dibromomethane	50.000	49.727	0.5	100	0.00
47 T	Bromodichloromethane	50.000	50.107	-0.2	98	0.00
48 T	Methyl methacrylate	50.000	45.485	9.0	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083122\
 Data File : VX030966.D
 Acq On : 31 Aug 2022 09:13
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 01 00:54:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 30 14:24:53 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	913.593	8.6	86	0.00
50 S	Toluene-d8	50.000	49.677	0.6	100	0.00
51 T	4-Methyl-2-Pentanone	250.000	220.199	11.9	83	0.00
52 CM	Toluene	50.000	50.880	-1.8#	100	0.00
53 T	t-1,3-Dichloropropene	50.000	52.505	-5.0	99	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.000	-2.0	97	0.00
55 T	1,1,2-Trichloroethane	50.000	49.839	0.3	101	0.00
56 T	Ethyl methacrylate	50.000	51.135	-2.3	97	0.00
57 T	1,3-Dichloropropane	50.000	49.611	0.8	99	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	250.737	-0.3	95	-0.01
59 T	2-Hexanone	250.000	236.087	5.6	91	0.00
60 T	Dibromochloromethane	50.000	52.803	-5.6	102	0.00
61 T	1,2-Dibromoethane	50.000	51.805	-3.6	99	0.00
62 S	4-Bromofluorobenzene	50.000	50.343	-0.7	100	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	96	0.00
64 T	Tetrachloroethene	50.000	56.293	-12.6	113	0.00
65 PM	Chlorobenzene	50.000	49.995	0.0	99	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	52.674	-5.3	103	0.00
67 C	Ethyl Benzene	50.000	51.271	-2.5#	101	0.00
68 T	m/p-Xylenes	100.000	105.402	-5.4	100	0.00
69 T	o-Xylene	50.000	52.112	-4.2	97	0.00
70 T	Styrene	50.000	51.736	-3.5	97	0.00
71 P	Bromoform	50.000	54.353	-8.7	99	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	95	0.00
73 T	Isopropylbenzene	50.000	51.049	-2.1	101	0.00
74 T	N-amyl acetate	50.000	46.419	7.2	91	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	46.535	6.9	96	0.00
76 T	1,2,3-Trichloropropane	50.000	50.021	-0.0	100	0.00
77 T	Bromobenzene	50.000	50.894	-1.8	104	0.00
78 T	n-propylbenzene	50.000	51.150	-2.3	100	0.00
79 T	2-Chlorotoluene	50.000	47.873	4.3	96	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.085	-2.2	99	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	50.802	-1.6	96	0.00
82 T	4-Chlorotoluene	50.000	50.702	-1.4	102	0.00
83 T	tert-Butylbenzene	50.000	49.806	0.4	96	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.425	-0.8	93	0.00
85 T	sec-Butylbenzene	50.000	51.909	-3.8	100	0.00
86 T	p-Isopropyltoluene	50.000	53.048	-6.1	104	0.00
87 T	1,3-Dichlorobenzene	50.000	54.018	-8.0	105	0.00
88 T	1,4-Dichlorobenzene	50.000	50.455	-0.9	102	0.00
89 T	n-Butylbenzene	50.000	53.788	-7.6	102	0.00
90 T	Hexachloroethane	50.000	54.025	-8.0	104	0.00
91 T	1,2-Dichlorobenzene	50.000	51.744	-3.5	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	48.298	3.4	89	0.00
93 T	1,2,4-Trichlorobenzene	50.000	53.520	-7.0	102	0.00
94 T	Hexachlorobutadiene	50.000	52.150	-4.3	103	0.00
95 T	Naphthalene	50.000	51.534	-3.1	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083122\
Data File : VX030966.D
Acq On : 31 Aug 2022 09:13
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Sep 01 00:54:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 30 14:24:53 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	52.691	-5.4	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6