



BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083122\
Data File : VX030965.D
Acq On : 31 Aug 2022 08:42
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
Title : SW846 8260
Last Update : Tue Aug 30 14:24:53 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.0	8121	PASS
75	95	30	60	48.5	24576	PASS
95	95	100	100	100.0	50661	PASS
96	95	5	9	6.9	3512	PASS
173	174	0.00	2	0.7	288	PASS
174	95	50	100	77.9	39461	PASS
175	174	5	9	7.4	2930	PASS
176	174	95	101	95.1	37544	PASS
177	176	5	9	6.7	2505	PASS