

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083123\
 Data File : VX037434.D
 Acq On : 31 Aug 2023 16:42
 Operator : JC/MD
 Sample : 04150-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE123D1-20230822

Quant Time: Sep 01 03:14:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	144093	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	236690	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	205018	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85839	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	88214	42.465	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	84.920%
35) Dibromofluoromethane	5.385	113	72347	46.153	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.300%
50) Toluene-d8	8.646	98	277213	48.753	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.500%
62) 4-Bromofluorobenzene	11.079	95	95402	42.694	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.380%
Target Compounds						
44) Trichloroethene	7.135	130	747	0.412	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083123\
Data File : VX037434.D
Acq On : 31 Aug 2023 16:42
Operator : JC/MD
Sample : 04150-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE123D1-20230822

Quant Time: Sep 01 03:14:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

