

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083123\
 Data File : VX037437.D
 Acq On : 31 Aug 2023 17:54
 Operator : JC/MD
 Sample : 04182-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Sep 01 03:15:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	142092	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	233723	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	206078	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92513	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	86644	42.296	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	84.600%
35) Dibromofluoromethane	5.385	113	71718	46.333	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.660%
50) Toluene-d8	8.647	98	272820	48.589	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.180%
62) 4-Bromofluorobenzene	11.079	95	98171	44.491	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083123\
Data File : VX037437.D
Acq On : 31 Aug 2023 17:54
Operator : JC/MD
Sample : 04182-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Sep 01 03:15:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

