

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083123\
 Data File : VX037445.D
 Acq On : 31 Aug 2023 21:08
 Operator : JC/MD
 Sample : 04251-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OW-03B-51.5-083023

Quant Time: Sep 01 03:17:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	144313	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	238384	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	214543	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95160	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	88256	42.420	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	84.840%
35) Dibromofluoromethane	5.385	113	72427	45.876	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.760%
50) Toluene-d8	8.647	98	281134	49.091	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.180%
62) 4-Bromofluorobenzene	11.079	95	101261	44.994	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083123\
Data File : VX037445.D
Acq On : 31 Aug 2023 21:08
Operator : JC/MD
Sample : 04251-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OW-03B-51.5-083023

Quant Time: Sep 01 03:17:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

