

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083123\
 Data File : VX037447.D
 Acq On : 31 Aug 2023 21:56
 Operator : JC/MD
 Sample : 04251-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OW-07B-65-083023-3V

Quant Time: Sep 01 03:18:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	144415	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	240081	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	214597	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96043	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	88918	42.708	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.420%
35) Dibromofluoromethane	5.385	113	74342	46.756	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.520%
50) Toluene-d8	8.647	98	283210	49.104	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.200%
62) 4-Bromofluorobenzene	11.079	95	101598	44.825	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.640%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.319	101	712	0.466	ug/l #	64
12) 1,1-Dichloroethene	2.319	96	120018	80.733	ug/l	83
16) Acetone	2.379	43	1921	2.359	ug/l #	87
21) trans-1,2-Dichloroethene	3.093	96	6705	3.968	ug/l #	82
24) 1,1-Dichloroethane	3.611	63	17821	6.019	ug/l	99
25) 2-Butanone	4.574	43	2742	2.236	ug/l #	88
27) cis-1,2-Dichloroethene	4.489	96	98753	49.568	ug/l	79
30) Chloroform	5.092	83	2673	0.804	ug/l	89
42) 1,2-Dichloroethane	6.086	62	10297	4.053	ug/l	95
44) Trichloroethene	7.123	130	106663	58.061	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083123\
Data File : VX037447.D
Acq On : 31 Aug 2023 21:56
Operator : JC/MD
Sample : 04251-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OW-07B-65-083023-3V

Quant Time: Sep 01 03:18:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

