

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083123\
 Data File : VX037451.D
 Acq On : 31 Aug 2023 23:33
 Operator : JC/MD
 Sample : 04252-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-04-12-083023

Quant Time: Sep 01 03:19:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	149444	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	250902	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	221537	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96384	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	94176	43.711	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.420%
35) Dibromofluoromethane	5.385	113	76329	45.935	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.880%
50) Toluene-d8	8.647	98	293913	48.762	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.520%
62) 4-Bromofluorobenzene	11.079	95	103293	43.607	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.220%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.325	101	16120	10.204	ug/l	96
44) Trichloroethene	7.123	130	152859	79.619	ug/l	95
64) Tetrachloroethene	9.275	164	1058	0.655	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083123\
Data File : VX037451.D
Acq On : 31 Aug 2023 23:33
Operator : JC/MD
Sample : 04252-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-04-12-083023

Quant Time: Sep 01 03:19:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

