



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083123\
Data File : VX037426.D
Acq On : 31 Aug 2023 13:14
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Title : SW846 8260
Last Update : Fri Aug 18 09:30:38 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	15.8	7606	PASS
75	95	30	60	48.6	23448	PASS
95	95	100	100	100.0	48227	PASS
96	95	5	9	7.1	3428	PASS
173	174	0.00	2	0.7	245	PASS
174	95	50	100	70.3	33907	PASS
175	174	5	9	7.5	2535	PASS
176	174	95	101	95.3	32328	PASS
177	176	5	9	6.3	2050	PASS