

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090119\
 Data File : VX012002.D
 Acq On : 01 Sep 2019 19:51
 Operator : SY/SP
 Sample : K4529-05
 Misc : 3.80g/5.0mL/MSVOA X/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T9-12-13-T2-(1.9)

Quant Time: Sep 02 05:20:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082919S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 30 01:59:05 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	396091	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	516017	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	391062	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	136725	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	164547	43.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.72%	
35) Dibromofluoromethane	5.48	113	150490	45.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.96%	
50) Toluene-d8	8.71	98	544718	46.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.10%	
62) 4-Bromofluorobenzene	11.14	95	159100	30.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	61.32%	
Target Compounds						
29) Tetrahydrofuran	5.12	42	4033	5.433	ug/l #	84


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090119\
Data File : VX012002.D
Acq On : 01 Sep 2019 19:51
Operator : SY/SP
Sample : K4529-05
Misc : 3.80µ/5.0mL/MSVOA X/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T9-12-13-T2-(1.9)

Quant Time: Sep 02 05:20:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082919S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 30 01:59:05 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.65 min Scan# 781

Delta R.T. 0.00 min

Lab File: VX012002.D

Acq: 01 Sep 2019 19:51

Instrument :

MSVOA_X

ClientSampleId :


T9-12-13-T2-(1.9)

Tot Ion:168 Resp: 396091

Ion Ratio Lower Upper

168 100

99 46.0 38.4 57.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.65



150000

100000

50000

0

Time--> 5.50 5.60 5.70 5.80



#29

Tetrahydrofuran

Concen: 5.433 ug/l

RT: 5.12 min Scan# 695

Delta R.T. 0.01 min

Lab File: VX012002.D

Acq: 01 Sep 2019 19:51


Tgt Ion: 42 Resp: 4033

Ion Ratio Lower Upper

42 100

72 51.4 36.2 54.4

71 27.0 33.1 49.7#



Abundance Ion 42.00 (41.70 to 42.70): VX0

Ion 72.00 (71.70 to 72.70): VX0

Ion 71.00 (70.70 to 71.70): VX0


1500


1000

500

0

Time--> 5.00 5.10 5.20









#33
 1,2-Dichloroethane-d4
 Concen: 43.360 ug/l
 RT: 6.05 min Scan# 847
 Delta R.T. 0.00 min
 Lab File: VX012002.D
 Acq: 01 Sep 2019 19:51




Instrument :
 MSVOA_X
 ClientSampleId :
 T9-12-13-T2-(1.9)


Tot Ion: 65 Resp: 164547
 Ion Ratio Lower Upper
 65 100
 67 50.1 0.0 102.2



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.85 min Scan# 978
 Delta R.T. 0.00 min
 Lab File: VX012002.D
 Acq: 01 Sep 2019 19:51

Tgt Ion: 114 Resp: 516017
 Ion Ratio Lower Upper
 114 100
 63 19.1 0.0 38.8
 88 15.6 0.0 31.2









#35
 Dibromofluoromethane
 Concen: 45.480 ug/l
 RT: 5.48 min Scan# 754
 Delta R.T. 0.00 min
 Lab File: VX012002.D
 Acq: 01 Sep 2019 19:51




Instrument :
 MSVOA_X
 ClientSampleId :
 T9-12-13-T2-(1.9)


Tot Ion:113 Resp: 150490
 Ion Ratio Lower Upper
 113 100
 111 102.8 82.1 123.1
 192 28.5 21.5 32.3



#50
 Toluene-d8
 Concen: 46.052 ug/l
 RT: 8.71 min Scan# 1283
 Delta R.T. 0.00 min
 Lab File: VX012002.D
 Acq: 01 Sep 2019 19:51

Tgt Ion: 98 Resp: 544718
 Ion Ratio Lower Upper
 98 100
 100 66.6 53.0 79.6





#62

4-Bromofluorobenzene

Concen: 30.663 ug/l

RT: 11.14 min Scan# 1681

Delta R.T. 0.00 min

Lab File: VX012002.D

Acq: 01 Sep 2019 19:51

Instrument :

MSVOA_X

ClientSampleId :

T9-12-13-T2-(1.9)


Tot Ion: 95 Resp: 159100

Ion Ratio Lower Upper

95 100

174 101.6 0.0 196.0




176 99.2 0.0 191.4



Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.11 min Scan# 1513

Delta R.T. 0.00 min

Lab File: VX012002.D

Acq: 01 Sep 2019 19:51


Tgt Ion: 117 Resp: 391062

Ion Ratio Lower Upper

117 100

82 49.9 42.0 63.0



119 32.5 26.2 39.4




Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): VX0

Ion 118.90 (118.60 to 119.60): V





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.07 min Scan# 1835

Delta R.T. 0.00 min

Lab File: VX012002.D

Acq: 01 Sep 2019 19:51

Instrument :

MSVOA_X

ClientSampleId :

T9-12-13-T2-(1.9)


Tot Ion:152 Resp: 136725

Ion Ratio Lower Upper

152 100

115 51.9 37.1 111.3

150 153.9 0.0 339.6



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

