

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090120\
 Data File : VX018184.D
 Acq On : 01 Sep 2020 12:13
 Operator : JC/SP
 Sample : PB131314TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB131314TB

Quant Time: Sep 02 03:38:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	481645	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	773521	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	751516	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	359153	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	335568	50.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.38%	
35) Dibromofluoromethane	5.47	113	265289	50.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.72%	
50) Toluene-d8	8.70	98	964780	49.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.34%	
62) 4-Bromofluorobenzene	11.12	95	364580	47.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.46%	
Target Compounds						
16) Acetone	2.42	43	33697	12.933	ug/l	95
18) Methyl Acetate	2.75	43	8873	1.411	ug/l	91
20) Methylene Chloride	2.83	84	11874	1.086	ug/l	97
43) Isopropyl Acetate	6.42	43	129607	9.684	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090120\
Data File : VX018184.D
Acq On : 01 Sep 2020 12:13
Operator : JC/SP
Sample : PB131314TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB131314TB

Quant Time: Sep 02 03:38:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

