

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090120\
 Data File : VX018188.D
 Acq On : 01 Sep 2020 13:45
 Operator : JC/SP
 Sample : PB131314ZHE#04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB131314ZHE#04

Quant Time: Sep 02 03:45:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	435945	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	715882	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	685726	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	322057	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	317342	52.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.88%	
35) Dibromofluoromethane	5.47	113	248563	50.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.96%	
50) Toluene-d8	8.70	98	899130	49.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.02%	
62) 4-Bromofluorobenzene	11.12	95	330899	46.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.64%	
Target Compounds						
16) Acetone	2.42	43	34540	14.647	ug/l	95
18) Methyl Acetate	2.75	43	7892	1.387	ug/l #	87
20) Methylene Chloride	2.83	84	11804	1.311	ug/l	94
43) Isopropyl Acetate	6.42	43	126327	10.199	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090120\
Data File : VX018188.D
Acq On : 01 Sep 2020 13:45
Operator : JC/SP
Sample : PB131314ZHE#04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB131314ZHE#04

Quant Time: Sep 02 03:45:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

