

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090120\
 Data File : VX018196.D
 Acq On : 01 Sep 2020 16:48
 Operator : JC/SP
 Sample : PB131314ZHE#12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB131314ZHE#12

Quant Time: Sep 02 04:03:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	421525	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	709428	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	705536	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	333529	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.04	65	319553	54.61	ug/l	0.00
Spiked Amount			Recovery	=	109.22%	
35) Dibromofluoromethane	5.46	113	247398	51.20	ug/l	0.00
Spiked Amount			Recovery	=	102.40%	
50) Toluene-d8	8.70	98	912010	50.68	ug/l	0.00
Spiked Amount			Recovery	=	101.36%	
62) 4-Bromofluorobenzene	11.12	95	340846	48.14	ug/l	0.00
Spiked Amount			Recovery	=	96.28%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	3.03	59	7757	6.817	ug/l	99
16) Acetone	2.42	43	34863	15.289	ug/l	94
18) Methyl Acetate	2.75	43	7553	1.373	ug/l #	74
20) Methylene Chloride	2.84	84	44012	8.470	ug/l	96
25) 2-Butanone	4.65	43	29325	7.825	ug/l	93
43) Isopropyl Acetate	6.41	43	146056	11.899	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090120\
Data File : VX018196.D
Acq On : 01 Sep 2020 16:48
Operator : JC/SP
Sample : PB131314ZHE#12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB131314ZHE#12

Quant Time: Sep 02 04:03:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

