

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090120\
 Data File : VX018199.D
 Acq On : 01 Sep 2020 17:57
 Operator : JC/SP
 Sample : PB131314ZHE#15
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB131314ZHE#15

Quant Time: Sep 02 04:10:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	418370	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	689748	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	685540	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	298362	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	312417	53.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.58%	
35) Dibromofluoromethane	5.46	113	240070	51.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.22%	
50) Toluene-d8	8.70	98	890437	50.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.78%	
62) 4-Bromofluorobenzene	11.12	95	318600	46.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.56%	
Target Compounds						
16) Acetone	2.42	43	34844	15.396	ug/l	92
18) Methyl Acetate	2.76	43	7713	1.412	ug/l #	82
20) Methylene Chloride	2.84	84	10800	1.194	ug/l	93
43) Isopropyl Acetate	6.42	43	125748	10.537	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090120\
Data File : VX018199.D
Acq On : 01 Sep 2020 17:57
Operator : JC/SP
Sample : PB131314ZHE#15
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB131314ZHE#15

Quant Time: Sep 02 04:10:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

