

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090120\
 Data File : VX018201.D
 Acq On : 01 Sep 2020 18:42
 Operator : JC/SP
 Sample : PB131314ZHE#17
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB131314ZHE#17

Quant Time: Sep 02 04:15:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	416769	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	680141	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	680468	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	301369	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	310838	53.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.44%	
35) Dibromofluoromethane	5.46	113	240189	51.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.70%	
50) Toluene-d8	8.70	98	863293	50.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.08%	
62) 4-Bromofluorobenzene	11.12	95	317584	46.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.58%	
Target Compounds						
16) Acetone	2.42	43	36958	16.393	ug/l #	80
18) Methyl Acetate	2.76	43	7673	1.410	ug/l #	85
20) Methylene Chloride	2.84	84	11016	1.251	ug/l	97
43) Isopropyl Acetate	6.42	43	123999	10.537	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090120\
Data File : VX018201.D
Acq On : 01 Sep 2020 18:42
Operator : JC/SP
Sample : PB131314ZHE#17
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB131314ZHE#17

Quant Time: Sep 02 04:15:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

