

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090120\
 Data File : VX018203.D
 Acq On : 01 Sep 2020 19:28
 Operator : JC/SP
 Sample : L3858-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HAM-WC-L-VB01

Manual Integrations
 APPROVED

MMDadoda
 9/2/2020 2:23:07 PM

Quant Time: Sep 02 07:35:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	423344	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	696725	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	663546	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	326214	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	315267	53.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.28%	
35) Dibromofluoromethane	5.46	113	248982	52.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.94%	
50) Toluene-d8	8.70	98	859799	48.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.30%	
62) 4-Bromofluorobenzene	11.12	95	315546	45.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.76%	
Target Compounds						
16) Acetone	2.42	43	19279	8.418	ug/l	94
30) Chloroform	5.18	83	4198m	0.485	ug/l	
52) Toluene	8.77	92	3984	0.333	ug/l	96
68) m/p-Xylenes	10.34	106	3589	0.427	ug/l	90
69) o-Xylene	10.68	106	2502	0.307	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090120\
Data File : VX018203.D
Acq On : 01 Sep 2020 19:28
Operator : JC/SP
Sample : L3858-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HAM-WC-L-VB01

Manual Integrations
APPROVED

MMDadoda
9/2/2020 2:23:07 PM

Quant Time: Sep 02 07:35:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

