

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX090120\
 Data File : VX018178.D
 Acq On : 01 Sep 2020 09:31
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 02 02:57:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	79	0.00
2 T	Dichlorodifluoromethane	0.609	0.654	-7.4	81	0.00
3 P	Chloromethane	0.642	0.615	4.2	77	0.00
4 C	Vinyl Chloride	0.639	0.613	4.1#	75	0.00
5 T	Bromomethane	0.308	0.364	-18.2	92	0.00
6 T	Chloroethane	0.343	0.365	-6.4	83	0.00
7 T	Trichlorofluoromethane	0.871	0.989	-13.5	89	0.00
8 T	Diethyl Ether	0.295	0.307	-4.1	83	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.471	0.493	-4.7	84	0.00
10 T	Methyl Iodide	0.438	0.575	-31.3#	96	0.00
11 T	Tert butyl alcohol	0.135	0.123	8.9	73	0.00
12 CM	1,1-Dichloroethene	0.489	0.464	5.1#	77	0.00
13 T	Acrolein	0.056	0.057	-1.8	82	0.00
14 T	Allyl chloride	0.878	0.892	-1.6	79	0.00
15 T	Acrylonitrile	0.306	0.286	6.5	75	0.00
16 T	Acetone	0.270	0.239	11.5	74	0.00
17 T	Carbon Disulfide	1.477	1.368	7.4	78	0.00
18 T	Methyl Acetate	0.653	0.672	-2.9	82	0.00
19 T	Methyl tert-butyl Ether	1.676	1.751	-4.5	82	0.00
20 T	Methylene Chloride	0.613	0.558	9.0	78	0.00
21 T	trans-1,2-Dichloroethene	0.541	0.541	0.0	82	0.00
22 T	Diisopropyl ether	1.733	1.739	-0.3	79	0.00
23 T	Vinyl Acetate	1.542	1.564	-1.4	77	0.00
24 P	1,1-Dichloroethane	0.986	0.988	-0.2	79	0.00
25 T	2-Butanone	0.445	0.393	11.7	71	0.00
26 T	2,2-Dichloropropane	0.803	0.962	-19.8	97	0.00
27 T	cis-1,2-Dichloroethene	0.624	0.612	1.9	81	0.00
28 T	Bromochloromethane	0.476	0.423	11.1	72	-0.01
29 T	Tetrahydrofuran	0.283	0.251	11.3	69	0.00
30 C	Chloroform	1.021	1.042	-2.1#	81	0.00
31 T	Cyclohexane	0.831	0.834	-0.4	79	0.00
32 T	1,1,1-Trichloroethane	0.937	0.991	-5.8	85	0.00
33 S	1,2-Dichloroethane-d4	0.694	0.661	4.8	79	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	74	0.00
35 S	Dibromofluoromethane	0.341	0.345	-1.2	76	-0.01
36 T	1,1-Dichloropropene	0.461	0.513	-11.3	81	0.00
37 T	Ethyl Acetate	0.536	0.523	2.4	73	-0.01
38 T	Carbon Tetrachloride	0.520	0.610	-17.3	85	0.00
39 T	Methylcyclohexane	0.504	0.572	-13.5	82	0.00
40 TM	Benzene	1.355	1.435	-5.9	78	0.00
41 T	Methacrylonitrile	0.294	0.281	4.4	72	-0.01
42 TM	1,2-Dichloroethane	0.519	0.587	-13.1	84	0.00
43 T	Isopropyl Acetate	0.865	0.871	-0.7	73	0.00
44 TM	Trichloroethene	0.383	0.415	-8.4	80	0.00
45 C	1,2-Dichloropropane	0.365	0.390	-6.8#	78	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX090120\
 Data File : VX018178.D
 Acq On : 01 Sep 2020 09:31
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 02 02:57:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.257	0.269	-4.7	78	0.00
47 T	Bromodichloromethane	0.513	0.574	-11.9	81	0.00
48 T	Methyl methacrylate	0.431	0.435	-0.9	74	0.00
49 T	1,4-Dioxane	0.008	0.008	0.0	71	0.00
50 S	Toluene-d8	1.268	1.241	2.1	73	0.00
51 T	4-Methyl-2-Pentanone	0.534	0.540	-1.1	72	0.00
52 CM	Toluene	0.859	0.930	-8.3#	79	0.00
53 T	t-1,3-Dichloropropene	0.568	0.640	-12.7	81	0.00
54 T	cis-1,3-Dichloropropene	0.602	0.660	-9.6	80	0.00
55 T	1,1,2-Trichloroethane	0.366	0.385	-5.2	77	0.00
56 T	Ethyl methacrylate	0.545	0.584	-7.2	76	0.00
57 T	1,3-Dichloropropane	0.604	0.629	-4.1	76	0.00
58 T	2-Chloroethyl Vinyl ether	0.245	0.237	3.3	71	0.00
59 T	2-Hexanone	0.406	0.423	-4.2	73	0.00
60 T	Dibromochloromethane	0.430	0.477	-10.9	80	0.00
61 T	1,2-Dibromoethane	0.398	0.419	-5.3	77	0.00
62 S	4-Bromofluorobenzene	0.499	0.514	-3.0	77	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	75	0.00
64 T	Tetrachloroethene	0.367	0.413	-12.5	84	0.00
65 PM	Chlorobenzene	0.999	1.057	-5.8	79	0.00
66 T	1,1,1,2-Tetrachloroethane	0.394	0.435	-10.4	81	0.00
67 C	Ethyl Benzene	1.698	1.863	-9.7#	79	0.00
68 T	m/p-Xylenes	0.634	0.721	-13.7	80	0.00
69 T	o-Xylene	0.614	0.679	-10.6	80	0.00
70 T	Styrene	1.046	1.189	-13.7	80	0.00
71 P	Bromoform	0.336	0.378	-12.5	82	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
73 T	Isopropylbenzene	3.298	3.538	-7.3	82	0.00
74 T	N-amyl acetate	1.565	1.511	3.5	72	0.00
75 P	1,1,2,2-Tetrachloroethane	1.163	1.043	10.3	74	0.00
76 T	1,2,3-Trichloropropane	1.112	1.047	5.8	77	0.00
77 T	Bromobenzene	0.912	0.919	-0.8	82	0.00
78 T	n-propylbenzene	3.745	4.102	-9.5	85	0.00
79 T	2-Chlorotoluene	2.314	2.469	-6.7	83	0.00
80 T	1,3,5-Trimethylbenzene	2.710	3.054	-12.7	84	0.00
81 T	trans-1,4-Dichloro-2-butene	0.412	0.410	0.5	79	0.00
82 T	4-Chlorotoluene	2.750	2.948	-7.2	84	0.00
83 T	tert-Butylbenzene	2.675	2.939	-9.9	84	0.00
84 T	1,2,4-Trimethylbenzene	2.771	3.094	-11.7	84	0.00
85 T	sec-Butylbenzene	3.042	3.459	-13.7	86	0.00
86 T	p-Isopropyltoluene	2.794	3.304	-18.3	88	0.00
87 T	1,3-Dichlorobenzene	1.601	1.693	-5.7	87	0.00
88 T	1,4-Dichlorobenzene	1.603	1.686	-5.2	86	0.00
89 T	n-Butylbenzene	2.550	2.902	-13.8	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX090120\
Data File : VX018178.D
Acq On : 01 Sep 2020 09:31
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Sep 02 02:57:06 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.550	0.630	-14.5	88	0.00
91 T	1,2-Dichlorobenzene	1.507	1.593	-5.7	83	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.291	0.269	7.6	76	0.00
93 T	1,2,4-Trichlorobenzene	1.005	1.093	-8.8	87	0.00
94 T	Hexachlorobutadiene	0.396	0.441	-11.4	91	0.00
95 T	Naphthalene	3.306	3.411	-3.2	79	0.00
96 T	1,2,3-Trichlorobenzene	0.993	1.092	-10.0	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6