

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX090120\
 Data File : VX018204.D
 Acq On : 01 Sep 2020 19:51
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 02 04:25:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	71	0.00
2 T	Dichlorodifluoromethane	50.000	50.539	-1.1	68	0.00
3 P	Chloromethane	50.000	47.302	5.4	68	0.00
4 C	Vinyl Chloride	50.000	46.429	7.1#	65	0.00
5 T	Bromomethane	50.000	57.744	-15.5	80	0.00
6 T	Chloroethane	50.000	50.908	-1.8	71	0.00
7 T	Trichlorofluoromethane	50.000	56.291	-12.6	79	0.00
8 T	Diethyl Ether	50.000	51.318	-2.6	73	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	52.248	-4.5	75	0.00
10 T	Methyl Iodide	50.000	53.702	-7.4	82	0.00
11 T	Tert butyl alcohol	250.000	253.663	-1.5	73	0.00
12 CM	1,1-Dichloroethene	50.000	49.059	1.9#	71	0.00
13 T	Acrolein	250.000	248.897	0.4	72	0.00
14 T	Allyl chloride	50.000	49.980	0.0	70	0.00
15 T	Acrylonitrile	250.000	252.697	-1.1	73	0.00
16 T	Acetone	250.000	253.229	-1.3	76	0.00
17 T	Carbon Disulfide	50.000	46.083	7.8	69	0.00
18 T	Methyl Acetate	50.000	56.244	-12.5	80	0.00
19 T	Methyl tert-butyl Ether	50.000	52.337	-4.7	74	0.00
20 T	Methylene Chloride	50.000	51.083	-2.2	71	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.651	0.7	73	0.00
22 T	Diisopropyl ether	50.000	50.753	-1.5	71	0.00
23 T	Vinyl Acetate	250.000	258.373	-3.3	71	0.00
24 P	1,1-Dichloroethane	50.000	50.620	-1.2	72	0.00
25 T	2-Butanone	250.000	244.330	2.3	70	0.00
26 T	2,2-Dichloropropane	50.000	46.492	7.0	67	0.00
27 T	cis-1,2-Dichloroethene	50.000	47.778	4.4	71	0.00
28 T	Bromochloromethane	50.000	51.463	-2.9	75	0.00
29 T	Tetrahydrofuran	250.000	243.942	2.4	68	0.00
30 C	Chloroform	50.000	51.839	-3.7#	74	0.00
31 T	Cyclohexane	50.000	48.693	2.6	69	0.00
32 T	1,1,1-Trichloroethane	50.000	52.130	-4.3	75	0.00
33 S	1,2-Dichloroethane-d4	50.000	51.428	-2.9	76	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	70	0.00
35 S	Dibromofluoromethane	50.000	51.270	-2.5	73	0.00
36 T	1,1-Dichloropropene	50.000	50.468	-0.9	69	0.00
37 T	Ethyl Acetate	50.000	49.700	0.6	70	0.00
38 T	Carbon Tetrachloride	50.000	55.612	-11.2	76	0.00
39 T	Methylcyclohexane	50.000	50.155	-0.3	68	0.00
40 TM	Benzene	50.000	50.420	-0.8	71	0.00
41 T	Methacrylonitrile	50.000	50.058	-0.1	71	0.00
42 TM	1,2-Dichloroethane	50.000	54.318	-8.6	76	0.00
43 T	Isopropyl Acetate	50.000	49.770	0.5	69	0.00
44 TM	Trichloroethene	50.000	50.135	-0.3	70	0.00
45 C	1,2-Dichloropropane	50.000	50.507	-1.0#	70	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX090120\
 Data File : VX018204.D
 Acq On : 01 Sep 2020 19:51
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 02 04:25:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
46 T	Dibromomethane	50.000	49.967	0.1	71	0.00
47 T	Bromodichloromethane	50.000	53.073	-6.1	73	0.00
48 T	Methyl methacrylate	50.000	50.154	-0.3	69	0.00
49 T	1,4-Dioxane	1000.000	1002.485	-0.2	65	0.00
50 S	Toluene-d8	50.000	49.002	2.0	69	0.00
51 T	4-Methyl-2-Pentanone	250.000	260.541	-4.2	70	0.00
52 CM	Toluene	50.000	51.201	-2.4#	71	0.00
53 T	t-1,3-Dichloropropene	50.000	51.303	-2.6	70	0.00
54 T	cis-1,3-Dichloropropene	50.000	49.810	0.4	69	0.00
55 T	1,1,2-Trichloroethane	50.000	51.159	-2.3	71	0.00
56 T	Ethyl methacrylate	50.000	52.655	-5.3	71	0.00
57 T	1,3-Dichloropropane	50.000	50.879	-1.8	70	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	236.482	5.4	65	0.00
59 T	2-Hexanone	250.000	263.927	-5.6	70	0.00
60 T	Dibromochloromethane	50.000	54.640	-9.3	75	0.00
61 T	1,2-Dibromoethane	50.000	51.397	-2.8	71	0.00
62 S	4-Bromofluorobenzene	50.000	51.617	-3.2	73	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	71	0.00
64 T	Tetrachloroethene	50.000	49.997	0.0	71	0.00
65 PM	Chlorobenzene	50.000	50.095	-0.2	70	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	52.291	-4.6	73	0.00
67 C	Ethyl Benzene	50.000	51.932	-3.9#	71	0.00
68 T	m/p-Xylenes	100.000	104.834	-4.8	70	0.00
69 T	o-Xylene	50.000	52.243	-4.5	71	0.00
70 T	Styrene	50.000	53.344	-6.7	71	0.00
71 P	Bromoform	50.000	55.348	-10.7	76	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	76	0.00
73 T	Isopropylbenzene	50.000	49.889	0.2	73	0.00
74 T	N-amyl acetate	50.000	47.429	5.1	67	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	45.913	8.2	72	0.00
76 T	1,2,3-Trichloropropane	50.000	45.269	9.5	70	0.00
77 T	Bromobenzene	50.000	48.309	3.4	75	0.00
78 T	n-propylbenzene	50.000	50.273	-0.5	74	0.00
79 T	2-Chlorotoluene	50.000	49.405	1.2	73	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.589	-3.2	73	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	45.923	8.2	69	0.00
82 T	4-Chlorotoluene	50.000	49.028	1.9	73	0.00
83 T	tert-Butylbenzene	50.000	52.121	-4.2	76	0.00
84 T	1,2,4-Trimethylbenzene	50.000	52.215	-4.4	75	0.00
85 T	sec-Butylbenzene	50.000	51.404	-2.8	74	0.00
86 T	p-Isopropyltoluene	50.000	53.156	-6.3	75	0.00
87 T	1,3-Dichlorobenzene	50.000	47.187	5.6	73	0.00
88 T	1,4-Dichlorobenzene	50.000	47.773	4.5	74	0.00
89 T	n-Butylbenzene	50.000	49.788	0.4	73	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX090120\
Data File : VX018204.D
Acq On : 01 Sep 2020 19:51
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Sep 02 04:25:28 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	50.000	53.477	-7.0	78	0.00
91 T	1,2-Dichlorobenzene	50.000	49.028	1.9	73	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	48.545	2.9	76	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.041	1.9	74	0.00
94 T	Hexachlorobutadiene	50.000	51.407	-2.8	80	0.00
95 T	Naphthalene	50.000	49.775	0.5	72	0.00
96 T	1,2,3-Trichlorobenzene	50.000	48.996	2.0	75	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6