

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090121\
 Data File : VX024059.D
 Acq On : 01 Sep 2021 12:40
 Operator : JC/MD
 Sample : M3610-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 YORK-ST-MH-1

Quant Time: Sep 02 07:08:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X080321W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Aug 03 12:05:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.910	128	31221	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.769	114	173344	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	155850	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.964	65	82488	34.164	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	113.867%#
60) 4-Bromofluorobenzene	11.085	95	70673	27.685	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	92.300%
63) Toluene-d8	8.653	98	209538	29.874	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.567%
Target Compounds						
15) Acetone	2.392	58	891	2.464	ug/l	Qvalue 76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090121\
Data File : VX024059.D
Acq On : 01 Sep 2021 12:40
Operator : JC/MD
Sample : M3610-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
YORK-ST-MH-1

Quant Time: Sep 02 07:08:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X080321W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Aug 03 12:05:42 2021
Response via : Initial Calibration

