

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090122\
 Data File : VX031018.D
 Acq On : 01 Sep 2022 10:33
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Sep 02 02:52:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	220979	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	370452	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	332207	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	157933	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	117472	41.315	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	82.640%
35) Dibromofluoromethane	5.385	113	113732	42.069	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	84.140%
50) Toluene-d8	8.653	98	424101	42.625	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	85.240%#
62) 4-Bromofluorobenzene	11.079	95	145098	41.289	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	82.580%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090122\
Data File : VX031018.D
Acq On : 01 Sep 2022 10:33
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Sep 02 02:52:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

