

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090122\
 Data File : VX031025.D
 Acq On : 01 Sep 2022 13:20
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Sep 02 02:54:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	226575	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	382929	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	329746	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	153863	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	118536	40.660	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	81.320%
35) Dibromofluoromethane	5.391	113	114413	40.942	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	81.880%
50) Toluene-d8	8.653	98	421361	40.970	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	81.940%#
62) 4-Bromofluorobenzene	11.079	95	151074	41.589	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	83.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090122\
Data File : VX031025.D
Acq On : 01 Sep 2022 13:20
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Sep 02 02:54:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

