

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090122\
 Data File : VX031033.D
 Acq On : 01 Sep 2022 20:32
 Operator : JC/MD
 Sample : N4281-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Sep 02 02:56:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	221563	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	356912	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	315629	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	136982	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	163164	57.234	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.460%
35) Dibromofluoromethane	5.391	113	155210	59.590	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	119.180%
50) Toluene-d8	8.653	98	558742	58.288	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	116.580%#
62) 4-Bromofluorobenzene	11.085	95	187069	55.252	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.500%
Target Compounds						
						Qvalue
44) Trichloroethene	7.129	130	31385	9.467	ug/l	78
64) Tetrachloroethene	9.281	164	1262	0.400	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090122\
Data File : VX031033.D
Acq On : 01 Sep 2022 20:32
Operator : JC/MD
Sample : N4281-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Sep 02 02:56:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

