

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090122\
 Data File : VX031035.D
 Acq On : 01 Sep 2022 21:19
 Operator : JC/MD
 Sample : N4281-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW201D1-20220819

Quant Time: Sep 02 02:56:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	219992	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	351468	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	306610	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	134050	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	155371	54.889	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.780%	
35) Dibromofluoromethane	5.391	113	150046	58.500	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 117.000%	
50) Toluene-d8	8.653	98	546760	57.921	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 115.840%#	
62) 4-Bromofluorobenzene	11.085	95	178428	53.516	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 107.040%	
Target Compounds						
44) Trichloroethene	7.129	130	5963	1.826	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090122\
Data File : VX031035.D
Acq On : 01 Sep 2022 21:19
Operator : JC/MD
Sample : N4281-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW201D1-20220819

Quant Time: Sep 02 02:56:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

